

Technical Evaluation Report								
Sr #	Tool Name	BUKHAT	PCM	TESCOL	MACS	MAKKAYS	AYK	UNICOM
1	Power Quality Analyzer	Not Quoted	Qualified	Qualified	Qualified	Qualified	Qualified	Qualified
2	Lux Meter	Qualified	Qualified	Qualified	Qualified	Qualified	Qualified	Qualified
3	IR Gun (Temperature Range 600 ° C)	Qualified	Qualified	Qualified	Qualified	Qualified	Qualified	Qualified
4	Hygrometer	Not Quoted	Qualified	Qualified	Qualified	Qualified	Qualified	Not Quoted
5	Air Flow Meters	Qualified	Qualified	Qualified	Qualified	Qualified	Not Quoted	Not Quoted
6	Anemometers	Not Quoted	Not Qualified	Not Qualified	Not Qualified	Not Qualified	Not Qualified	Not Quoted
7	Thermal Imager A	Qualified	Qualified	Qualified	Qualified	Qualified	Not Qualified	Qualified
	Thermal Imager B	Qualified	Not Qualified	Qualified	Qualified	Qualified	Not Qualified	Qualified
8	Power Logger (3 Phase)	Not Quoted	Qualified	Qualified	Qualified	Qualified	Qualified	Qualified
9	Portable Air Flow Meter for Compressors	Not Quoted	Qualified	Not Qualified	Not Quoted	Not Quoted	Not Quoted	Not Quoted
10	Flue Gas Analyzer	Not Quoted	Qualified	Not Quoted	Not Qualified	Not Quoted	Not Quoted	Not Quoted
11	Air Quality Meter	Qualified	Qualified	Not Qualified	Not Quoted	Qualified	Not Quoted	Not Quoted

Tools (Not Qualified)	Bidders	Reasons
Anemometer	M/S PCM, TESCOL, MACS, MAKKAYS, AYK	All qualified bidders quoted tools are below velocity range as per tender specifications
Thermal Imager A & B	M/S AYK	M/S AYK quoted specification below for both cases as per tender specifications
Portable Air Flow Meter For Compressor	M/S TESCOL	Parameters does not comply with tender specifications
Air Quality Meter	M/S TESCOL	Quoted tool only measure PM and no other parameter compliance.